



UL 61010-2-032

STANDARD FOR SAFETY

Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-032: Particular Requirements for Hand-Held and Hand-Manipulated Current Sensors for Electrical Test and Measurement

ULNORM.COM: Click to view the full PDF of UL 61010-2-032 2024

ULNORM.COM : Click to view the full PDF of UL 61010-2-032 2024

UL Standard for Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-032: Particular Requirements for Hand-Held and Hand-Manipulated Current Sensors for Electrical Test and Measurement, UL 61010-2-032

Third Edition, Dated October 17, 2024

Summary of Topics

This new Third Edition of ANSI/UL 61010-2-032 dated October 17, 2024 is an Adoption of IEC 61010-2-032, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-032: Particular Requirements for Hand-Held and Hand-Manipulated Current Sensors for Electrical Test and Measurement (fifth edition, issued by IEC September 2023) as a new IEC-based UL standard with US National Differences.

The requirements are substantially in accordance with Proposal(s) on this subject dated August 2, 2024.

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted in any form by any means, electronic, mechanical photocopying, recording, or otherwise without prior permission of ULSE Inc. (ULSE).

ULSE provides this Standard "as is" without warranty of any kind, either expressed or implied, including but not limited to, the implied warranties of merchantability or fitness for any purpose.

In no event will ULSE be liable for any special, incidental, consequential, indirect or similar damages, including loss of profits, lost savings, loss of data, or any other damages arising out of the use of or the inability to use this Standard, even if ULSE or an authorized ULSE representative has been advised of the possibility of such damage. In no event shall ULSE's liability for any damage ever exceed the price paid for this Standard, regardless of the form of the claim.

Users of the electronic versions of UL's Standards for Safety agree to defend, indemnify, and hold ULSE harmless from and against any loss, expense, liability, damage, claim, or judgment (including reasonable attorney's fees) resulting from any error or deviation introduced while purchaser is storing an electronic Standard on the purchaser's computer system.

No Text on This Page

ULNORM.COM : Click to view the full PDF of UL 61010-2-032 2024

OCTOBER 17, 2024



ANSI/UL 61010-2-032-2024

1

UL 61010-2-032

**Standard for Safety Requirements for Electrical Equipment for
Measurement, Control, and Laboratory Use –
Part 2-032: Particular Requirements for Hand-Held and Hand-Manipulated
Current Sensors for
Electrical Test and Measurement**

First Edition – August, 2014
Second Edition – January, 2020

Third Edition

October 17, 2024

This ANSI/UL Standard for Safety consists of the Third Edition.

The most recent designation of ANSI/UL 61010-2-032 as an American National Standard (ANSI) occurred on October 17, 2024. ANSI approval for a standard does not include the Cover Page, Transmittal Pages, Title Page, or Preface. The National Difference Page and IEC Foreword are also excluded from the ANSI approval of IEC-based standards.

Comments or proposals for revisions on any part of the Standard may be submitted to ULSE at any time. Proposals should be submitted via a Proposal Request in the Collaborative Standards Development System (CSDS) at <https://csds.ul.com>.

Our Standards for Safety are copyrighted by ULSE Inc. Neither a printed nor electronic copy of a Standard should be altered in any way. All of our Standards and all copyrights, ownerships, and rights regarding those Standards shall remain the sole and exclusive property of ULSE Inc.

© 2024 ULSE Inc. All rights reserved.

No Text on This Page

ULNORM.COM : Click to view the full PDF of UL 61010-2-032 2024

CONTENTS

PREFACE (UL)	7
NATIONAL DIFFERENCES	9
FOREWORD	11
INTRODUCTION	15
1 Scope and object	17
1.1.1 Equipment included in scope	17
1.1.2 Equipment excluded from scope	17
1.2.1 Aspects included in scope	17
2 Normative references	18
2DV Modification: Add the following	19
3 Terms and definitions.....	19
3.1 Equipment and states of equipment	19
3.2 Parts and accessories.....	19
3.5 Safety terms.....	19
3.6 Insulation	20
4 Tests	20
4.3.2.5 Power supply.....	20
4.3.2.6 Input and output voltages or currents.....	20
4.4.2.8 Outputs	21
4.4.2.101 Surge protective devices.....	21
5 Marking and documentation	21
5.1.2 Identification.....	21
5.1.5 TERMINALS, connections and operating devices	22
5.4.1 General.....	24
5.4.3 Equipment installation	24
5.4.4 Equipment operation.....	24
6 Protection against electric shock	25
6.1.2 Exceptions	25
6.5.1 General.....	25
6.5.2 Not used	26
6.5.5 Not used	26
6.6 Connections to external circuits	26
6.7.1.3 Creepage distances	30
6.7.1.5 Requirements for insulation according to type of circuit	30
6.8.1 General.....	31
6.8.3.1 The a.c. voltage test.....	31
6.8.3.2 The d.c. voltage test.....	31
6.101 Protection against HAZARDOUS LIVE conductors	31
7 Protection against mechanical HAZARDS	34
8 Resistance to mechanical stresses.....	34
8.101 JAW ENDS abrasion test	34
8.102 JAW impact test	35
8.103 Pressure test at high temperature for insulation of flexible current sensors	36
8.104 Pull test for endcaps of flexible current sensors	38
8.105 Attachment of the input/output leads.....	39
9 Protection against the spread of fire and arc flash	42
9.101 Protection of measuring circuits	42
9.102 Prevention of HAZARD from arc flash and short-circuits	47
10 Equipment temperature limits and resistance to heat.....	49

10.5	Resistance to heat	49
10.101	Other temperatures of current sensors	50
11	Protection against HAZARDS from fluids and solid foreign objects	50
12	Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	50
13	Protection against liberated gases and substances, explosion and implosion	50
14	Components and subassemblies	50
14.101	Probe assemblies and accessories	51
15	Protection by interlocks	51
16	HAZARDS resulting from application	51
17	RISK assessment	51
101	Measuring circuits	51
101.1	General	51
101.2	Current sensor with an internal current transformer	51
101.3	Indicating devices	52

Annexes

Annex D (normative) Parts between which insulation requirements are specified(see 6.4 and 6.5.3)

Annex F (normative) Routine tests

F.1	General	57
F.101	JAWS OF current sensors	57

Annex K (normative) Insulation requirements not covered by 6.7

K.2	Insulation in secondary circuits	58
K.2.1	General	58
K.3	Insulation in circuits not addressed in 6.7, Clause K.1, Clause K.2 or Clause K.101 , and for measuring circuits where measurement categories do not apply	58
K.3.1	General	58
K.3.2	CLEARANCE calculation	59
K.3.101	CLEARANCES between MAINS CIRCUITS and input/output circuits	62
K.4	Attenuation of TRANSIENT OVERVOLTAGE levels	64
K.101	Insulation requirements for measuring circuits rated for measurement categories	66
K.101.1	General	66
K.101.2	CLEARANCES	66
K.101.3	CREEPAGE DISTANCES	67
K.101.4	Solid insulation	67

Annex L (informative) Index of defined terms

Annex AA (informative) MEASUREMENT CATEGORIES

AA.1	General	73
AA.2	MEASUREMENT CATEGORIES	73
AA.2.1	MEASUREMENT CATEGORY II	73
AA.2.2	MEASUREMENT CATEGORY III	73
AA.2.3	MEASUREMENT CATEGORY IV	73
AA.2.4	Measuring circuits without a MEASUREMENT CATEGORY RATING	73

Annex BB (informative) HAZARDS pertaining to measurements performed in certain environments

BB.1	General.....	77
BB.2	HAZARDS.....	77
	BB.2.1 General.....	77
	BB.2.2 Electric shock.....	77
	BB.2.3 Arc flash and arc blast.....	77
	BB.2.4 Thermal burns.....	77
BB.3	Telecommunications networks.....	78
BB.4	Current measurements in inductive circuits.....	78
BB.5	Battery-driven circuits.....	78
BB.6	Measurements at higher frequencies.....	78
BB.7	Measurements using measuring circuits with a FUNCTIONAL EARTH TERMINAL.....	78

Annex CC (informative) 4 mm "banana" TERMINALS

CC.1	General.....	80
CC.2	Dimensions.....	80

Annex DD (informative) Flowchart for insulation according to the type of circuit

Annex EE (normative) Determination of CLEARANCES for [Table 101](#)

Annex FF (normative) Type of current sensors

Annex GG (normative) CLAMP MULTIMETER

GG.1	General.....	86
GG.2	Current sensor type.....	87
GG.3	CLAMP MULTIMETER RATING.....	87
GG.4	Marking of measuring circuit TERMINALS and JAWS.....	87
GG.5	Probe assemblies and accessories.....	87

Bibliography

No Text on This Page

[ULNORM.COM](https://ulnorm.com) : Click to view the full PDF of UL 61010-2-032 2024

PREFACE (UL)

This UL Standard is based on IEC Publication 61010-2-032: fifth edition Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement. IEC publication 61010-2-032 is copyrighted by the IEC.

This edition has been issued to satisfy ULSE Standards policy.

This UL Standard 61010-2-032 Standard for Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-032: Particular Requirements for Hand-Held and Hand-Manipulated Current Sensors for Electrical Test and Measurement, is to be used in conjunction with the latest edition of UL 61010-1. The requirements for measurement equipment for insulation resistance and test equipment for electric strength are contained in this Part 2 Standard and UL 61010-1.

Requirements of this Part 2 Standard, where stated, amend the requirements of UL 61010-1.

Where a particular subclause of UL 61010-1 is not mentioned in UL 61010-2-032, the UL 61010-1 subclause applies.

These materials are subject to copyright claims of IEC and ULSE. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of ULSE. All requests pertaining to the Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-032: Particular Requirements for Hand-Held and Hand-Manipulated Current Sensors for Electrical Test and Measurement, UL 61010-2-032 Standard should be submitted to ULSE.

Note - Although the intended primary application of this Standard is stated in its Scope, it is important to note that it remains the responsibility of the users of the Standard to judge its suitability for their particular purpose.

ULNORM.COM : Click to view the full PDF of UL 61010-2-032-032

No Text on This Page

ULNORM.COM : Click to view the full PDF of UL 61010-2-032 2024

NATIONAL DIFFERENCES

National Differences from the text of International Electrotechnical Commission (IEC) Publication 61010-2-032, Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement, copyright 2023 are indicated by notations (differences) and are presented in bold text.

There are five types of National Differences as noted below. The difference type is noted on the first line of the National Difference in the standard. The standard may not include all types of these National Differences.

D1 – These are National Differences which are based on **basic safety principles and requirements**, elimination of which would compromise safety for consumers and users of products.

D2 – These are National Differences from IEC requirements based on existing **safety practices**. These requirements reflect national safety practices, where empirical substantiation (for the IEC or national requirement) is not available or the text has not been included in the IEC standard.

DC – These are National Differences based on the **component standards** and will not be deleted until a particular component standard is harmonized with the IEC component standard.

DE – These are National Differences based on **editorial comments or corrections**.

DR – These are National Differences based on the **national regulatory requirements**.

Each national difference contains a description of what the national difference entails. Typically one of the following words is used to explain how the text of the national difference is to be applied to the base IEC text:

Addition / Add - An addition entails adding a complete new numbered clause, subclause, table, figure, or annex. Addition is not meant to include adding select words to the base IEC text.

Modification / Modify - A modification is an altering of the existing base IEC text such as the addition, replacement or deletion of certain words or the replacement of an entire clause, subclause, table, figure, or annex of the base IEC text.

Deletion / Delete - A deletion entails complete deletion of an entire numbered clause, subclause, table, figure, or annex without any replacement text.

No Text on This Page

ULNORM.COM : Click to view the full PDF of UL 61010-2-032 2024

FOREWORD

INTERNATIONAL ELECTROTECHNICAL COMMISSION

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE – Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.

2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.

3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.

4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.

5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.

6) All users should ensure that they have the latest edition of this publication.

7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.

8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.

9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61010-2-032 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment. It is an International Standard.

This fifth edition cancels and replaces the fourth edition published in 2019. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) In 1.1.1, definitions of current sensor types have been moved to a new Annex FF;
- b) Clause 2, all normative references have been dated and new normative references have been added;

- c) 3.2.103, a new definition PROTECTIVE FINGERGUARD has been added which replaces the previous definition of PROTECTIVE BARRIER;
- d) 4.4.2.101 is a new subclause about surge protective devices;
- e) in 5.1.5.101.2, minimum RATINGS for voltage of measuring TERMINALS are required;
- f) Subclause 6.5.1 has been modified;
- g) Subclause 6.5.5 is no longer used;
- h) Subclause 6.6.101 modifies 6.6.101 and 6.6.102 of previous edition:
- 1) in 6.6.101.1, insulating material of group I may be allowed for determination of CREEPAGE DISTANCES of measuring circuit TERMINALS;
 - 2) in 6.6.101.2, CLEARANCES and CREEPAGE DISTANCES up to 3 000 V for measuring circuit TERMINALS in unmated position have been defined;
 - 3) in 6.6.101.3, requirements for measuring circuit TERMINALS in partially mated position have been specified;
 - 4) in 6.6.101.4, requirements for measuring circuit TERMINALS in mated position have been specified;
 - 5) Subclause 6.6.101.5 replaces 6.6.102;
- i) Subclause 6.6.102 replaces 6.101 of previous edition with modifications;
- j) Subclause 6.101.2 replaces 6.9.101.1 of previous edition with modifications;
- k) Subclause 6.101.3 replaces 6.9.101.2 of previous edition with modifications;
- l) Subclause 6.101.4 replaces 6.9.102 of previous edition with modifications;
- m) in 8.101, JAW ENDS abrasion test has been modified;
- n) 8.105 is a new subclause for input/output leads attachment has been added;
- o) in 9.101.2, relocation of 101.3 of previous edition;
- p) in 9.101.3, relocation of 101.4 of previous edition, extension to MEASUREMENT CATEGORY II and reference to IEC 61000-4-5 for tests;
- q) Table 104 has been replaced by Table K.101;
- r) in 9.102, relocation of Clause 102 of previous edition;
- s) in 14.101, relocation of 14.102. Subclause 14.101 of previous edition has been deleted;
- t) 101.3 is a new subclause for protections against HAZARD occurring from reading a voltage value in replacement of Clause EE.5 of previous edition;
- u) in Table D.101, transients are disregarded for insulation between JAW ENDS and input/output circuits;

- v) in Clause F.101, test voltages for routine test of JAWS have been modified;
- w) in K.2.1, another method for determination of CLEARANCES of secondary circuits is proposed;
- x) in K.3.2, new Table K.15 and Table K.16 for CLEARANCE calculation;
- y) K.3.101 is a new clause;
- z) Clause K.4, redraft of the clause to propose a method for determination of U_t for circuits which reduce TRANSIENT OVERVOLTAGES;
- aa) Table K.101 replaces Table 104;
- bb) Subclause K.101.4 has been reviewed and tables and tests for solid insulation have been modified;
- cc) Table K.104 of previous edition has been deleted;
- dd) Annex AA: Figure AA.1 has been redesigned;
- ee) Annex EE: addition of a new informative annex for determination of CLEARANCES for Table 101;
- ff) Annex GG: this annex was Annex EE of previous edition and the current sensor type of a CLAMP MULTIMETER is type A or type B.

The text of this International Standard is based on the following documents:

Draft	Report on voting
66/788A/FDIS	66/798/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

A list of all parts of the IEC 61010 series, under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

This document is to be used in conjunction with IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016.

This document supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement*.

Where a particular subclause of IEC 61010-1 is not mentioned in this document, that subclause applies as far as is reasonable. Where this document states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in IEC 61010-1 should be adapted accordingly.

In this standard:

a) the following print types are used:

- requirements: in roman type;
- NOTES: in small roman type;
- *conformity and tests: in italic type;*
- terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS;

b) subclauses, figures, tables and notes which are additional to those in IEC 61010-1 are numbered starting from 101. Additional annexes are lettered starting from AA and additional list items are lettered from aa).

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

101DV DE *Modification to add the following to IEC Foreword:*

The numbering system in the standard uses a space instead of a comma to indicate thousands and uses a comma instead of a period to indicate a decimal point. For example, 1 000 means 1,000 and 1,01 means 1.01.

102DV DE *Modification to add the following to the IEC Foreword:*

For this Standard, all references to "IEC 61010-1:2010" or "IEC 61010 1:2010/AMD1:2016" refer to UL 61010-1.

For this Standard, all references to IEC 61010-2-030, IEC 61010-2-033 and IEC 61010-2-034 refer to UL versions of the particulars.

For this Standard, all references to IEC 61010-031 refer to UL 61010-031.